

A Simple Improvement on Existing Technology to Enhance Transseptal Puncture and Left Atrial Access Using Large Diameter Sheaths for Mapping, Ablation and LAA Occlusion Procedures

Gregory K. Feld, MD, Jonathan C. Hsu, MD, MAS

From the University of California San Diego Health System, Department of Medicine,
Division of Cardiology

Running Title: Large Diameter Dilator to Assist Transseptal Puncture with Large Sheaths

Funding Source: None

Editorial Comment

Word Count: 870

Key Words: Watchman™, transseptal puncture, left atrial appendage, atrial fibrillation, ablation

Address for Correspondence:

9452 Medical Center Drive

MC7411, Altman Bldg, 3rd Floor, Room 3E-313

La Jolla, CA 92037

Disclosures:

Dr. Feld, as Director of the UCSD EP Fellowship Training program, has received fellowship stipend support from Boston Scientific, Biotronik, Biosense Webster, Medtronic, and Abbott Medical, is co-founder and co-owner of Perminova, has received consulting fees from Acutus Medical, Vektor Medical, and Irysis Pharmaceuticals, and has equity interest in Medwaves and Acutus Medical.

Dr. Hsu has received honoraria from Medtronic, Abbott, Boston Scientific, Biotronik, Biosense-Webster, Zoll Medical, Janssen Pharmaceuticals, Bristol-Myers Squibb, and Pfizer, has research grants with Biotronik and Biosense-Webster, and has equity interest in Acutus Medical and Vektor Medical.

Since the earliest description of transseptal catheterization techniques by Ross, et.al. in the late 1950's, to the modern era of cardiac electrophysiology which relies heavily on the same technique, the evolution of technology has been critical to transseptal catheterization improvement, including ease of performance and safety¹⁻³. In this issue of the Journal of Cardiac Electrophysiology⁴, Inohara, et.al. report their benchtop research and first in human experience with a new transseptal dilator (ExpanSure Large Access Transseptal Dilator®, Baylis Medical, Montreal, Canada) used to aid left atrial (LA) access for large sheaths commonly employed in procedures such as the Watchman™ LA appendage occluder deployment.

The authors employed a standard test system (Instron Testing System, Norwood, MA) to measure force and displacement in-vitro while advancing the new ExpanSure® dilator compared to a standard 8.5F Swartz™ SL1 sheath and dilator (Abbott Medical, St. Paul, MN) through a model silicone septum. In addition, during in-vivo studies in man, they determined time to positioning of a Watchman™ delivery sheath (Boston Scientific, St. Paul, MN) in the LA during a standard LA appendage occlusion procedure and administered a survey to the operators to subjectively assess crossing force and tissue resistance observed after predilating with the Expansure® dilator.

The authors found during their in-vitro studies that use of the Expansure® dilator reduced peak crossing force and total work by 20% and 38%, respectively, compared to a standard SL1 sheath when advanced through the silicone septum, and a 12% reduction in peak crossing force and 20% reduction in total work for the Watchman™ sheath when advanced through the silicone septum after pre-dilatation with the Expansure® dilator compared to the SL1 sheath. They also found during their in-vivo studies in patients that prior to deployment of the Watchman™ device in the LA appendage, delivery of the initial large guiding sheath to the LA

was accomplished in a short time, and subjectively with a low crossing force and tissue resistance in most cases. However, there was no statistical comparison of these objective or subjective variables during the Watchman™ procedures performed following pre-dilation with the Expansure® dilator to those following pre-dilation with a standard SL1 sheath or without pre-dilation.

While the use of a large dilator to transition from a standard transseptal crossing tool (e.g. 8.5 French SL1 sheath) to one of several large diameter transseptal sheaths commonly in use today (e.g. Watchman™ guide sheath 14 French OD, Flexcath Medtronic guide sheaths 13 or 15 French OD, Acutus guide sheath 16 French OD) is obviously logical to allow for greater ease in crossing the inter-atrial septum, until now this technology has not been commercially available⁵⁻⁷. Theoretical approaches for septal dilatation after initial puncture have been proposed using balloons, stents, and expanding sheaths, but these approaches would involve use of additional complex equipment and be time consuming. A simple approach such as that described in this paper using the Expansure® dilator alone with a transseptal needle (i.e. without the need for another standard 8.5 French sheath system first followed by a dilator exchange) is technically appealing, potentially cost saving, and if the results are reproducible, may enhance delivery of a large diameter sheath into the LA. Transseptal catheterization can be difficult in contemporary practice (particularly in a repeat procedure where the septum may be stiff due to fibrosis from previous healing of the access point) when transitioning from a standard size sheath such as an 8.5 French transseptal sheath to a large diameter 13-16 French sheath for mapping, ablation, or LAA occluder device delivery. Perhaps the only potential concern if using the Expansure® dilator alone versus a standard 8.5 French SL1 sheath would be the larger profile of the system when performing initial transseptal catheterization, and the maneuverability of this larger system

in achieving the desired transseptal access site. However, this did not appear to be a concern from subjective operator responses during this study.

In summary, while randomized clinical trials comparing the Expansure® dilator to a standard transseptal sheath for inter-atrial septal crossing prior to upsizing to a large diameter sheath for ease and safety would provide definitive evidence supporting its use, the data presented in this paper, in addition to the logic behind this approach and low probability of any increased risk, make it reasonable to adopt in clinical practice without mandating further investigation.

References

1. Ross J Jr. Transseptal left heart catheterization: a new method of left atrial puncture. *Ann Surg* 1959;149:395-401.
2. Ross J Jr. Transseptal left heart catheterization: A 50-year odyssey. *J Am Coll Cardiol* 2008;51:2107-15.
3. Feld GK, Tiongson J, Oshodi G. Particle formation and risk of embolization during transseptal catheterization: comparison of standard transseptal needles and a new radiofrequency transseptal needle. *J Interven Card Electrophysiol* 2011;30:31-36.
4. Inohara T, Gilhofer T, Al-Dujaili S, Leung L, Yeung D, Tsang M, Saw J. A new transseptal solution for enabling left atrial access of large delivery sheaths. *J Card Electrophysiol* 2021;??:??-??
5. Holmes D, Doshi S, Kar S, Price MJ, Sanchez JM, Sievert H, Valderrabano M, Reddy VY J. Left Atrial Appendage Closure as an Alternative to Warfarin for Stroke Prevention in Atrial Fibrillation: A Patient-Level Meta-Analysis. *J Am Coll Cardiol* 2015;65:2624-7.
6. Ströker E, De Greef Y, Schwagten B, Kupics K, Coutiño HE, Takarada K, Abugattas JP, Salghetti F, De Cocker J, Stockman D, Sieira J, Brugada P, Chierchia GB, de Asmundis C. Over-the-needle trans-septal access using the cryoballoon delivery sheath and dilator in atrial fibrillation ablation. *Pacing Clin Electrophysiol* 2019;42:868-873.
7. Shi R, Parikh P, Chen Z, Angel N, Norman M, Hussain W, Butcher C, Haldar S, Jones DG, Riad O, Markides V, Wong T. Validation of Dipole Density Mapping During Atrial Fibrillation and Sinus Rhythm in Human Left Atrium. *JACC Clin Electrophysiol* 2020;6:182-184.